

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE	Calib. No.:	00033
Calibration File No.:	00633	Cert. Date:	09/22/2016
Certification File No.:	00634	Lin. Date:	09/22/2016
Linearity File No.:	00635	Soln. Date:	10/08/2016
Solution File No.:	00638	File Date:	10/08/2016
Sequential File No.:	00638		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	15D138	Expires:	04/13/2017
		Bottle No.:	3700

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:47D		
Control 1 EC	0.097%	21:47D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.098%	21:47D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:48D		
Control 2 EC	0.097%	21:49D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:49D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:49D		
Control 3 EC	0.097%	21:50D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.098%	21:50D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BISE

First Name: JOHN

MI: E

Signature: Sgt. John Bise + 60

Badge No.: 60

Date: 10/08/2016

