

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00633	Calib. Date: 09/22/2016
Certification File No.:	00634	Calib. No.: 00033
Linearity File No.:	00635	Cert. Date: 09/22/2016
Solution File No.:	00640	Cert. No.: 00024
Sequential File No.:	00640	Lin. Date: 09/22/2016
		Lin. No.: 00028
		Soln. Date: 11/05/2016
		Soln. No.: 00147
		File Date: 11/05/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0162
Solution Control Lot:	15D138	Expires: 04/13/2017
		Bottle No.: 3580

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:38S		
Control 1 EC	0.098%	21:39S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:39S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:40S		
Control 2 EC	0.098%	21:40S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.098%	21:40S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:41S		
Control 3 EC	0.097%	21:42S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.098%	21:42S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:42S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BISE

First Name: JOHN

MI: E

Signature: \_\_\_\_\_

*Sgt. John Bise #60*

Badge No.: 60

Date: 11/05/2016

NOV 07 2016

NOV 07 2016