

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00481 Calib. Date: 04/10/2014 Calib. No.: 00023
Certification File No.: 00482 Cert. Date: 04/10/2014 Cert. No.: 00015
Linearity File No.: 00483 Lin. Date: 04/10/2014 Lin. No.: 00018
Solution File No.: 00489 Soln. Date: 05/09/2014 Soln. No.: 00110
Sequential File No.: 00489 File Date: 05/09/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 11/12/2015
Solution Control Lot: 13K125 Bottle No.: 1227

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:44D		
Control 1 EC	0.097%	09:45D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:45D		
Control 2 EC	0.098%	09:46D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:46D		
Control 3 EC	0.098%	09:47D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:47D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Received By
PATROL
MAY 13 2014

Changed By:

Last Name: UNGRADY

First Name: EDWARD

MI: K

Signature: _____

Badge No.: 099

Date: 05/09/2014

