

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0069  
 Location: MARLBORO TWP. POLICE  
 Calibration File No.: 00633 Calib. Date: 09/22/2016 Calib. No.: 00033  
 Certification File No.: 00634 Cert. Date: 09/22/2016 Cert. No.: 00024  
 Linearity File No.: 00635 Lin. Date: 09/22/2016 Lin. No.: 00028  
 Solution File No.: 00636 Soln. Date: 09/22/2016 Soln. No.: 00145  
 Sequential File No.: 00636 File Date: 09/22/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162  
 Control Solution %: 0.100% Expires: 04/13/2017  
 Solution Control Lot: 15D138 Bottle No.: 0365

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:29D		
Control 1 EC	0.099%	10:29D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:29D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:30D		
Control 2 EC	0.098%	10:30D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:30D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:31D		
Control 3 EC	0.098%	10:32D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:32D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:32D		


All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DD UJ P2 - 066 (DK)

**Changed By:**

Last Name: KLIMIK First Name: DAVID MI: W

Signature:  #7040 Badge No.: 7040  
Date: 09/22/2016