

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0069  
Location: MARLBORO TWP. POLICE  
Calibration File No.: 00706 Calib. Date: 01/09/2018 Calib. No.: 00037  
Certification File No.: 00707 Cert. Date: 01/09/2018 Cert. No.: 00028  
Linearity File No.: 00708 Lin. Date: 01/09/2018 Lin. No.: 00032  
Solution File No.: 00713 Soln. Date: 02/23/2018 Soln. No.: 00176  
Sequential File No.: 00713 File Date: 02/23/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162  
Control Solution %: 0.100% Expires: 09/14/2019  
Solution Control Lot: 17290 Bottle No.: 0603

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:49S		
Control 1 EC	0.099%	10:50S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:51S		
Control 2 EC	0.097%	10:51S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:52S		
Control 3 EC	0.098%	10:53S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:54S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: RUDITSKY

First Name: DAVID

MI: A

Signature: \_\_\_\_\_



Badge No.: 090

Date: 02/23/2018

