

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00681 Calib. Date: 07/27/2017 Calib. No.: 00036
Certification File No.: 00682 Cert. Date: 07/27/2017 Cert. No.: 00027
Linearity File No.: 00683 Lin. Date: 07/27/2017 Lin. No.: 00031
Solution File No.: 00685 Soln. Date: 08/16/2017 Soln. No.: 00166
Sequential File No.: 00685 File Date: 08/16/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0164
Control Solution %: 0.100% Expires: 01/30/2019
Solution Control Lot: 17050 Bottle No.: 0945

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:40D		
Control 1 EC	0.103%	09:41D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.102%	09:41D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:41D		
Control 2 EC	0.103%	09:42D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.103%	09:42D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:43D		
Control 3 EC	0.103%	09:43D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.103%	09:43D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: LOYER

First Name: JOHN

MI: B

Signature: _____



Badge No.: 057

Date: 08/16/2017

