

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00582	Calib. Date:	10/16/2015
Certification File No.:	00583	Cert. Date:	10/16/2015
Linearity File No.:	00584	Lin. Date:	10/16/2015
Solution File No.:	00595	Soln. Date:	01/12/2016
Sequential File No.:	00595	File Date:	01/12/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	15D138	Expires:	04/13/2017
		Bottle No.:	0362

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:11S		
Control 1 EC	0.099%	10:11S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:11S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:12S		
Control 2 EC	0.098%	10:13S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.097%	10:13S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:13S		
Control 3 EC	0.098%	10:14S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:14S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TISCHLER

First Name: MICHAEL

MI: J

Signature: St. M. Tischler 075

Badge No.: 075

Date: 01/12/2016