

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00582 Calib. Date: 10/16/2015 Calib. No.: 00030
Certification File No.: 00583 Cert. Date: 10/16/2015 Cert. No.: 00021
Linearity File No.: 00584 Lin. Date: 10/16/2015 Lin. No.: 00025
Solution File No.: 00585 Soln. Date: 10/16/2015 Soln. No.: 00132
Sequential File No.: 00585 File Date: 10/16/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 04/13/2017
Solution Control Lot: 15D138 Bottle No.: 0369

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:54D		
Control 1 EC	0.102%	14:54D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:54D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:55D		
Control 2 EC	0.101%	14:56D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:56D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:56D		
Control 3 EC	0.101%	14:57D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.102%	14:57D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DD4JPZ - 066 KA

Changed By:

Last Name: ALCOTT First Name: KEVIN MI: W.
Signature: TPR. I [Signature] #6704 Badge No.: 6704
Date: 10/16/2015