

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE	Calib. No.:	00024
Calibration File No.:	00512	Cert. Date:	10/09/2014
Certification File No.:	00513	Lin. Date:	10/09/2014
Linearity File No.:	00515	Soln. Date:	11/08/2014
Solution File No.:	00521	File Date:	11/08/2014
Sequential File No.:	00521		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	13K125	Expires:	11/12/2015
		Bottle No.:	1242

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:23S		
Control 1 EC	0.101%	21:24S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:24S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:24S		
Control 2 EC	0.100%	21:25S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:26S		
Control 3 EC	0.100%	21:26S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:26S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:27S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ARMINIO

First Name: GERALD


MI: A

Signature: _____

[Handwritten Signature] #080

Badge No.: 080

Date: 11/08/2014



DEPARTMENT OF
Traffic and Public Safety
This is to certify that


Gerald A. Arminio
Marlboro Township Police Department

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE ALCOTEST 7110 MKIII-C
A METHOD TO DETERMINE INTOXICATION
GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY, THIS 14th DAY OF August

TWO THOUSAND AND Eight

Gerald A. Arminio
SUPERINTENDENT
NEW JERSEY STATE POLICE

Paul J. ...
ATTORNEY GENERAL
STATE OF NEW JERSEY




ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1. 1-19-10	OCRA	Wm ...
2. 08-09-12	OCRA	F. ...
3. 1/16/14	LAKEHURST	Adam ...
4.		
5.		
6.		
7.		
8.		
9.		

DEPARTMENT OF

Traffic and Public Safety



This is to certify that

Gerald A. Armino

Marlboro Twp. Police Department

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE ALCOHOL TEST 7110 MKH C A METHOD TO DETERMINE INTOXICATION

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 14th DAY OF August

TWO THOUSAND AND EIGHT

[Signature]
SUPERINTENDENT
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY