

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE	Calib. No.:	00033
Calibration File No.:	00633	Cert. Date:	09/22/2016
Certification File No.:	00634	Lin. Date:	09/22/2016
Linearity File No.:	00635	Soln. Date:	12/21/2016
Solution File No.:	00648	File Date:	12/21/2016
Sequential File No.:	00648		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	15D138	Expires:	04/13/2017
		Bottle No.:	0357

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:33S		
Control 1 EC	0.099%	23:34S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:34S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:34S		
Control 2 EC	0.097%	23:35S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:35S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:36S		
Control 3 EC	0.098%	23:36S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.098%	23:36S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:37S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ARMINIO

First Name: GERALD

MI: A

Signature: _____



Badge No.: 080

Date: 12/21/2016

DEC 28 2016