

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00657	Calib. Date:	02/02/2017
Certification File No.:	00658	Cert. No.:	00034
Linearity File No.:	00659	Lin. Date:	02/02/2017
Solution File No.:	00662	Cert. No.:	00025
Sequential File No.:	00662	Lin. No.:	00029
		Soln. Date:	03/17/2017
		Soln. No.:	00155
		File Date:	03/17/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	201605D	Expires:	05/17/2018
		Bottle No.:	0205

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:14S		
Control 1 EC	0.102%	07:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:16S		
Control 2 EC	0.101%	07:16S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:17S		
Control 3 EC	0.100%	07:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: LOYER

First Name: JOHN

MI: B

Signature: _____



Badge No.: 057

Date: 03/17/2017

MAR 23 2017