

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00543	Calib. Date: 03/23/2015
Certification File No.:	00544	Calib. No.: 00025
Linearity File No.:	00545	Cert. Date: 03/23/2015
Solution File No.:	00546	Cert. No.: 00017
Sequential File No.:	00546	Lin. Date: 03/23/2015
		Lin. No.: 00021
		Soln. Date: 03/23/2015
		Soln. No.: 00122
		File Date: 03/23/2015

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0164
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1229

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:12S	17.12 D (1b)	
Control 1 EC	0.102%	16:13S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:14S		
Control 2 EC	0.102%	16:14S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:15S		
Control 3 EC	0.102%	16:16S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:17S	17.17 D (1b)	

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDVJ P2-080 (1b)

**Changed By:**

Last Name: DENNIS

First Name: MARC

MI: W.

Signature: SGT. [Signature] #5925

Badge No.: 5925

Date: 03/23/2015

MAR 24 2015