

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00608	Calib. Date:	03/23/2016
Certification File No.:	00609	Cert. No.:	00032
Linearity File No.:	00610	Lin. Date:	03/23/2016
Solution File No.:	00611	Cert. No.:	00023
Sequential File No.:	00611	Lin. No.:	00027
		Soln. No.:	00138
		File Date:	03/23/2016

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0164
Control Solution %:	0.100%			Expires:	04/13/2017
Solution Control Lot:	15D138			Bottle No.:	0364

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:07S	12:07 D (PK)	
Control 1 EC	0.101%	11:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:08S		
Control 2 EC	0.100%	11:09S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:09S		
Control 3 EC	0.100%	11:10S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:11S	12:11 D (PK)	

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJ P2 - 080 (PK)

Changed By:

Last Name: KLIMIK

First Name: DAVID

MI: W

Signature: David D. Klimik #7040

Badge No.: 7040

Date: 03/23/2016