

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARXA-0069 |
| Location: | MARLBORO TWP. POLICE | |
| Calibration File No.: | 00608 | Calib. Date: 03/23/2016 |
| Certification File No.: | 00609 | Calib. No.: 00032 |
| Linearity File No.: | 00610 | Cert. Date: 03/23/2016 |
| Solution File No.: | 00611 | Cert. No.: 00023 |
| Sequential File No.: | 00611 | Lin. Date: 03/23/2016 |
| | | Lin. No.: 00027 |
| | | Soln. Date: 03/23/2016 |
| | | Soln. No.: 00138 |
| | | File Date: 03/23/2016 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDXC S3-0164 |
| Solution Control Lot: | 15D138 | Expires: 04/13/2017 |
| | | Bottle No.: 0364 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 11:07S | 12:07 D (PK) | |
| Control 1 EC | 0.101% | 11:07S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 11:07S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:08S | | |
| Control 2 EC | 0.100% | 11:09S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.101% | 11:09S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:09S | | |
| Control 3 EC | 0.100% | 11:10S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.101% | 11:10S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:11S | 12:11 D (PK) | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJ P2 - 080 (PK)

Changed By:

Last Name: KLIMIK

First Name: DAVID

MI: W

Signature: David D. Klimik #7040

Badge No.: 7040

Date: 03/23/2016