

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
 Location: MARLBORO TWP. POLICE
 Calibration File No.: 00512 Calib. Date: 10/09/2014 Calib. No.: 00024
 Certification File No.: 00513 Cert. Date: 10/09/2014 Cert. No.: 00016
 Linearity File No.: 00515 Lin. Date: 10/09/2014 Lin. No.: 00020
 Solution File No.: 00537 Soln. Date: 03/05/2015 Soln. No.: 00121
 Sequential File No.: 00537 File Date: 03/05/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0164
 Control Solution %: 0.100% Expires: 11/12/2015
 Solution Control Lot: 13K125 Bottle No.: 1234

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:47S		
Control 1 EC	0.100%	03:47S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	03:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:48S		
Control 2 EC	0.099%	03:49S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	03:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:49S		
Control 3 EC	0.099%	03:50S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:51S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ARMINIO

First Name: GERALD

MI: A

Signature: _____

[Handwritten Signature] #080

Badge No.: 80

Date: 03/05/2015

MAR 05 2015