

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00608 Calib. Date: 03/23/2016 Calib. No.: 00032
Certification File No.: 00609 Cert. Date: 03/23/2016 Cert. No.: 00023
Linearity File No.: 00610 Lin. Date: 03/23/2016 Lin. No.: 00027
Solution File No.: 00618 Soln. Date: 05/10/2016 Soln. No.: 00140
Sequential File No.: 00618 File Date: 05/10/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0164
Control Solution %: 0.100% Expires: 04/13/2017
Solution Control Lot: 15D138 Bottle No.: 0374

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:29D		
Control 1 EC	0.100%	14:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:31D		
Control 2 EC	0.100%	14:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:32D		
Control 3 EC	0.099%	14:32D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TISCHLER

First Name: MICHAEL

MI: J

Signature: Sgt. M. Tischler

Badge No.: 075

Date: 05/10/2016

