

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARXA-0069 |
| Location: | MARLBORO TWP. POLICE | |
| Calibration File No.: | 00657 | Calib. Date: 02/02/2017 |
| Certification File No.: | 00658 | Cert. No.: 00034 |
| Linearity File No.: | 00659 | Cert. Date: 02/02/2017 |
| Solution File No.: | 00669 | Lin. No.: 00029 |
| Sequential File No.: | 00669 | Lin. Date: 02/02/2017 |
| | | Soln. No.: 00159 |
| | | Soln. Date: 05/12/2017 |
| | | File Date: 05/12/2017 |

| | | | | | |
|-----------------------|---------|------------|-------|-------------|--------------|
| Calibrating Unit: | WET | Model No.: | CU-34 | Serial No.: | DDXC S3-0162 |
| Control Solution %: | 0.100% | | | Expires: | 05/17/2018 |
| Solution Control Lot: | 201605D | | | Bottle No.: | 0206 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 14:40D | | |
| Control 1 EC | 0.103% | 14:41D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.102% | 14:41D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:41D | | |
| Control 2 EC | 0.102% | 14:42D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.103% | 14:42D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:43D | | |
| Control 3 EC | 0.102% | 14:43D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.103% | 14:43D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:44D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: _____

Lt. B. #96

Badge No.: 96

Date: 05/12/2017

