

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0069  
Location: MARLBORO TWP. POLICE  
Calibration File No.: 00543 Calib. Date: 03/23/2015 Calib. No.: 00025  
Certification File No.: 00544 Cert. Date: 03/23/2015 Cert. No.: 00017  
Linearity File No.: 00545 Lin. Date: 03/23/2015 Lin. No.: 00021  
Solution File No.: 00554 Soln. Date: 06/18/2015 Soln. No.: 00125  
Sequential File No.: 00554 File Date: 06/18/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0164  
Control Solution %: 0.100% Expires: 11/12/2015  
Solution Control Lot: 13K125 Bottle No.: 1230

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:24D		
Control 1 EC	0.100%	09:24D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:25D		
Control 2 EC	0.100%	09:26D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:26D		
Control 3 EC	0.100%	09:27D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: Lt. *Bryan* #096

Badge No.: 96

Date: 06/18/2015

