

copy

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
 Location: MARLBORO TWP. POLICE
 Calibration File No.: 00608 Calib. Date: 03/23/2016 Calib. No.: 00032
 Certification File No.: 00609 Cert. Date: 03/23/2016 Cert. No.: 00023
 Linearity File No.: 00610 Lin. Date: 03/23/2016 Lin. No.: 00027
 Solution File No.: 00627 Soln. Date: 08/02/2016 Soln. No.: 00143
 Sequential File No.: 00627 File Date: 08/02/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0164
 Control Solution %: 0.100% Expires: 04/13/2017
 Solution Control Lot: 15D138 Bottle No.: 0368

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:00D		
Control 1 EC	0.099%	10:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:01D		
Control 2 EC	0.098%	10:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:03D		
Control 3 EC	0.099%	10:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
 Last Name: TISCHLER First Name: MICHAEL MI: J

Signature: Sgt. M. J. [Signature] 075 Badge No.: 075
 Date: 08/02/2016

