

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE	Calib. No.:	00019
Calibration File No.:	00413	Calib. Date:	04/04/2013
Certification File No.:	00414	Cert. Date:	04/04/2013
Linearity File No.:	00415	Lin. Date:	04/04/2013
Solution File No.:	00440	Soln. Date:	08/25/2013
Sequential File No.:	00440	File Date:	08/25/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	12F102	Expires:	06/28/2014
		Bottle No.:	0390

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:34D		
Control 1 EC	0.097%	11:35D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	11:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:35D		
Control 2 EC	0.096%	11:36D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:37D		
Control 3 EC	0.096%	11:37D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:38D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: \_\_\_\_\_

*Sgt. Bme #016*

Badge No.: 96

Date: 08/25/2013

