

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARXA-0069 |
| Location: | MARLBORO TWP. POLICE | |
| Calibration File No.: | 00481 | Calib. Date: 04/10/2014 |
| Certification File No.: | 00482 | Calib. No.: 00023 |
| Linearity File No.: | 00483 | Cert. Date: 04/10/2014 |
| Solution File No.: | 00510 | Cert. No.: 00015 |
| Sequential File No.: | 00510 | Lin. Date: 04/10/2014 |
| | | Lin. No.: 00018 |
| | | Soln. Date: 10/05/2014 |
| | | Soln. No.: 00115 |
| | | File Date: 10/05/2014 |

| | | | | | |
|-----------------------|--------|------------|-------|-------------|--------------|
| Calibrating Unit: | WET | Model No.: | CU-34 | Serial No.: | DDXC S3-0162 |
| Control Solution %: | 0.100% | | | Expires: | 11/12/2015 |
| Solution Control Lot: | 13K125 | | | Bottle No.: | 1228 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 22:01D | | |
| Control 1 EC | 0.096% | 22:01D | 34.1°C | *** TEST PASSED *** |
| Control 1 IR | 0.097% | 22:01D | 34.1°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:02D | | |
| Control 2 EC | 0.096% | 22:02D | 34.1°C | *** TEST PASSED *** |
| Control 2 IR | 0.097% | 22:02D | 34.1°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:03D | | |
| Control 3 EC | 0.096% | 22:04D | 34.1°C | *** TEST PASSED *** |
| Control 3 IR | 0.097% | 22:04D | 34.1°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 22:04D | | |

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: Lt.  #096

Badge No.: 96

Date: 10/05/2014

OCT 06 2014