

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00481	Calib. Date: 04/10/2014
Certification File No.:	00482	Calib. No.: 00023
Linearity File No.:	00483	Cert. Date: 04/10/2014
Solution File No.:	00510	Cert. No.: 00015
Sequential File No.:	00510	Lin. Date: 04/10/2014
		Lin. No.: 00018
		Soln. Date: 10/05/2014
		Soln. No.: 00115
		File Date: 10/05/2014

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0162
Control Solution %:	0.100%			Expires:	11/12/2015
Solution Control Lot:	13K125			Bottle No.:	1228

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:01D		
Control 1 EC	0.096%	22:01D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.097%	22:01D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:02D		
Control 2 EC	0.096%	22:02D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.097%	22:02D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:03D		
Control 3 EC	0.096%	22:04D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.097%	22:04D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:04D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: Lt.  #096

Badge No.: 96

Date: 10/05/2014

OCT 06 2014