

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00481 Calib. Date: 04/10/2014 Calib. No.: 00023
Certification File No.: 00482 Cert. Date: 04/10/2014 Cert. No.: 00015
Linearity File No.: 00483 Lin. Date: 04/10/2014 Lin. No.: 00018
Solution File No.: 00498 Soln. Date: 06/27/2014 Soln. No.: 00112
Sequential File No.: 00498 File Date: 06/27/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 11/12/2015
Solution Control Lot: 13K125 Bottle No.: 1232

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:10D		
Control 1 EC	0.096%	22:11D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	22:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:12D		
Control 2 EC	0.096%	22:12D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.097%	22:12D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:13D		
Control 3 EC	0.096%	22:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	22:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: _____

Sgt. Am #96

Badge No.: 96

Date: 06/27/2014

