

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00481	Calib. Date: 04/10/2014
Certification File No.:	00482	Calib. No.: 00023
Linearity File No.:	00483	Cert. Date: 04/10/2014
Solution File No.:	00502	Cert. No.: 00015
Sequential File No.:	00502	Lin. Date: 04/10/2014
		Lin. No.: 00018
		Soln. Date: 07/30/2014
		Soln. No.: 00113
		File Date: 07/30/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0162
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1241

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:20D		
Control 1 EC	0.097%	09:21D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:21D		
Control 2 EC	0.097%	09:22D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:23D		
Control 3 EC	0.097%	09:23D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: \_\_\_\_\_

*SA. [Signature]* #096

Badge No.: 96

Date: 07/30/2014

