

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00445	Calib. Date: 09/19/2013
Certification File No.:	00446	Cert. No.: 00020
Linearity File No.:	00447	Cert. Date: 09/19/2013
Solution File No.:	00475	Lin. No.: 00013
Sequential File No.:	00475	Lin. Date: 09/19/2013
		Soln. No.: 00107
		Soln. Date: 03/12/2014
		File Date: 03/12/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0164
Solution Control Lot:	12F102	Expires: 06/28/2014
		Bottle No.: 0400

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:01S		
Control 1 EC	0.100%	10:02S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:02S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:03S		
Control 2 EC	0.098%	10:04S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:05S		
Control 3 EC	0.099%	10:05S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:05S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:06S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: \_\_\_\_\_

*Sgt. [Signature] #2016*

Badge No.: 96

Date: 03/12/2014

