

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0069  
Location: MARLBORO TWP. POLICE  
Calibration File No.: 00445 Calib. Date: 09/19/2013 Calib. No.: 00020  
Certification File No.: 00446 Cert. Date: 09/19/2013 Cert. No.: 00013  
Linearity File No.: 00447 Lin. Date: 09/19/2013 Lin. No.: 00016  
Solution File No.: 00452 Soln. Date: 10/20/2013 Soln. No.: 00098  
Sequential File No.: 00452 File Date: 10/20/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0164  
Control Solution %: 0.100% Expires: 06/28/2014  
Solution Control Lot: 12F102 Bottle No.: 0393

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:57D		
Control 1 EC	0.099%	03:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	03:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:58D		
Control 2 EC	0.096%	03:59D	34.0°C	AMBIENT AIR CHECK ERR
Control 2 IR	-.---%	03:59D	34.0°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	-.---%	03:59D		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LAYMAN

First Name: JOSHUA

MI: E

Signature: \_\_\_\_\_

Badge No.: 101

Date: 10/20/2013

# Calibrating Unit

## New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C  
Location: MARLBORO TWP. POLICE  
Serial No.: ARXA-0069  
Calibration File No.: 00445 Calib. Date: 09/19/2013 Calib. No.: 00020  
Certification File No.: 00446 Cert. Date: 09/19/2013 Cert. No.: 00013  
Linearity File No.: 00447 Lin. Date: 09/19/2013 Lin. No.: 00016  
Solution File No.: 00453 Soln. Date: 10/23/2013 Soln. No.: 00099  
Sequential File No.: 00453 File Date: 10/23/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0164  
Control Solution %: 0.100% Expires: 06/28/2014  
Solution Control Lot: 12F102 Bottle No.: 0392

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:53D		
Control 1 EC	0.099%	08:53D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:54D		
Control 2 EC	0.097%	08:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:56D		
Control 3 EC	0.098%	08:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: \_\_\_\_\_



Badge No.: 96

Date: 10/23/2013