

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE	Calib. No.:	00019
Calibration File No.:	00413	Cert. Date:	04/04/2013
Certification File No.:	00414	Lin. Date:	04/04/2013
Linearity File No.:	00415	Soln. Date:	08/01/2013
Solution File No.:	00434	File Date:	08/01/2013
Sequential File No.:	00434		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	12F102	Expires:	06/28/2014
		Bottle No.:	0386

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:50D		
Control 1 EC	0.097%	09:50D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	09:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:51D		
Control 2 EC	0.096%	09:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	09:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:52D		
Control 3 EC	0.096%	09:53D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: _____

Sgt. [Signature] #0916

Badge No.: 96

Date: 08/01/2013

