

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00582 Calib. Date: 10/16/2015 Calib. No.: 00030
Certification File No.: 00583 Cert. Date: 10/16/2015 Cert. No.: 00021
Linearity File No.: 00584 Lin. Date: 10/16/2015 Lin. No.: 00025
Solution File No.: 00598 Soln. Date: 02/10/2016 Soln. No.: 00136
Sequential File No.: 00598 File Date: 02/10/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 04/13/2017
Solution Control Lot: 15D138 Bottle No.: 0372

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:42S		
Control 1 EC	0.100%	18:43S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	18:43S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:44S		
Control 2 EC	0.099%	18:44S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.101%	18:44S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:45S		
Control 3 EC	0.100%	18:46S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:46S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:46S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: _____

lt. [Signature] #096

Badge No.: 96

Date: 02/10/2016

FEB 16 2016