

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|-----------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARXA-0069 |
| Location: | MARLBORO TWP. POLICE | |
| Calibration File No.: | 00413 | Calib. Date: 04/04/2013 |
| Certification File No.: | 00414 | Calib. No.: 00019 |
| Linearity File No.: | 00415 | Cert. Date: 04/04/2013 |
| Solution File No.: | 00430 | Cert. No.: 00012 |
| Sequential File No.: | 00430 | Lin. Date: 04/04/2013 |
| | | Lin. No.: 00015 |
| | | Soln. Date: 07/04/2013 |
| | | Soln. No.: 00093 |
| | | File Date: 07/04/2013 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDXC S3-0162 |
| Solution Control Lot: | 12F102 | Expires: 06/28/2014 |
| | | Bottle No.: 0385 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 07:54D | | |
| Control 1 EC | 0.097% | 07:55D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.098% | 07:55D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:55D | | |
| Control 2 EC | 0.096% | 07:56D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.098% | 07:56D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:57D | | |
| Control 3 EC | 0.097% | 07:57D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.097% | 07:57D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:58D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: _____

Sgt. [Signature] #096

Badge No.: 96

Date: 07/04/2013

