

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00445	Calib. Date: 09/19/2013
Certification File No.:	00446	Calib. No.: 00020
Linearity File No.:	00447	Cert. Date: 09/19/2013
Solution File No.:	00458	Cert. No.: 00013
Sequential File No.:	00458	Lin. Date: 09/19/2013
		Lin. No.: 00016
		Soln. Date: 11/21/2013
		Soln. No.: 00100
		File Date: 11/21/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0164
Solution Control Lot:	12F102	Expires: 06/28/2014
		Bottle No.: 0397

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:57S		
Control 1 EC	0.100%	09:57S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58S		
Control 2 EC	0.098%	09:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:00S		
Control 3 EC	0.098%	10:00S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: _____

[Handwritten Signature] #096

Badge No.: 96

Date: 11/21/2013

