

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE	Calib. No.:	00029
Calibration File No.:	00577	Calib. Date:	10/07/2015
Certification File No.:	00578	Cert. Date:	10/07/2015
Linearity File No.:	00579	Lin. Date:	10/07/2015
Solution File No.:	00581	Soln. Date:	10/07/2015
Sequential File No.:	00581	File Date:	10/07/2015

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXC S3-0162
Control Solution %:	0.100%	Expires:	11/12/2015	Bottle No.:	1250
Solution Control Lot:	13K125				

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:39D		
Control 1 EC	0.100%	13:40D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:40D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:40D		
Control 2 EC	0.100%	13:41D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:41D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:42D		
Control 3 EC	0.100%	13:42D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:42D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJ P2-066

**Changed By:**

Last Name: DENNIS

First Name: MARC

MI: W.

Signature: SGT [Signature] #5925

Badge No.: 5925

Date: 10/07/2015