

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00543	Calib. Date:	03/23/2015
Certification File No.:	00544	Cert. Date:	03/23/2015
Linearity File No.:	00545	Lin. Date:	03/23/2015
Solution File No.:	00551	Soln. Date:	05/21/2015
Sequential File No.:	00551	File Date:	05/21/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0164
Solution Control Lot:	13K125	Expires:	11/12/2015
		Bottle No.:	1235

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:02D		
Control 1 EC	0.099%	23:03D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	23:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:03D		
Control 2 EC	0.099%	23:04D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	23:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:05D		
Control 3 EC	0.099%	23:05D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	23:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: WILLETT

First Name: BRIAN

MI: W

Signature: _____



Badge No.: 110

Date: 05/21/2015

