

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00512	Calib. Date:	10/09/2014
Certification File No.:	00513	Cert. Date:	10/09/2014
Linearity File No.:	00515	Lin. Date:	10/09/2014
Solution File No.:	00523	Soln. Date:	12/08/2014
Sequential File No.:	00523	File Date:	12/08/2014
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	13K125	Expires:	11/12/2015
		Bottle No.:	1247

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:00S		
Control 1 EC	0.098%	17:01S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:02S		
Control 2 EC	0.098%	17:02S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	17:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:03S		
Control 3 EC	0.099%	17:04S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:04S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: \_\_\_\_\_

*Ut. [Signature] #096*

Badge No.: 96

Date: 12/08/2014

