

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00543	Calib. Date: 03/23/2015
Certification File No.:	00544	Cert. No.: 00025
Linearity File No.:	00545	Cert. Date: 03/23/2015
Solution File No.:	00565	Lin. Date: 03/23/2015
Sequential File No.:	00565	Lin. No.: 00021
		Soln. Date: 08/21/2015
		Soln. No.: 00128
		File Date: 08/21/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0164
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1245

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:30D		
Control 1 EC	0.099%	19:31D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:32D		
Control 2 EC	0.099%	19:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	19:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:33D		
Control 3 EC	0.100%	19:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	19:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SNYDER

First Name: TIMOTHY

MI: J

Signature: _____

[Handwritten Signature] # 105

Badge No.: 105

Date: 08/21/2015

