

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00413	Calib. Date:	04/04/2013
Certification File No.:	00414	Cert. No.:	00019
Linearity File No.:	00415	Cert. Date:	04/04/2013
Solution File No.:	00425	Lin. No.:	00015
Sequential File No.:	00425	Soln. Date:	05/31/2013
		Soln. No.:	00091
		File Date:	05/31/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	12F102	Expires:	06/28/2014
		Bottle No.:	0381

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:51D		
Control 1 EC	0.098%	01:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	01:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:52D		
Control 2 EC	0.097%	01:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	01:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:54D		
Control 3 EC	0.097%	01:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	01:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: _____



Badge No.: 96

Date: 05/31/2013

