

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00633	Calib. Date: 09/22/2016
Certification File No.:	00634	Calib. No.: 00033
Linearity File No.:	00635	Cert. Date: 09/22/2016
Solution File No.:	00638	Cert. No.: 00024
Sequential File No.:	00638	Lin. Date: 09/22/2016
		Lin. No.: 00028
		Soln. Date: 10/08/2016
		Soln. No.: 00146
		File Date: 10/08/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0162
Solution Control Lot:	15D138	Expires: 04/13/2017
		Bottle No.: 3700

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:47D		
Control 1 EC	0.097%	21:47D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.098%	21:47D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:48D		
Control 2 EC	0.097%	21:49D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:49D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:49D		
Control 3 EC	0.097%	21:50D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.098%	21:50D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BISE

First Name: JOHN

MI: E

Signature: _____

Sgt. John Bise + 60

Badge No.: 60

Date: 10/08/2016

