

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE	Calib. No.:	00039
Calibration File No.:	00767	Cert. Date:	12/06/2018
Certification File No.:	00768	Lin. Date:	12/06/2018
Linearity File No.:	00769	Soln. Date:	02/04/2019
Solution File No.:	00778	File Date:	02/04/2019
Sequential File No.:	00778		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	0038

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:04S		
Control 1 EC	0.101%	12:04S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	12:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:05S		
Control 2 EC	0.099%	12:06S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:07S		
Control 3 EC	0.100%	12:07S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: CHAPLINSKI

First Name: JOSEPH

MI: S

Signature: _____



Badge No.: 61

Date: 02/04/2019

