

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00965 Calib. Date: 12/20/2021 Calib. No.: 00048
Certification File No.: 00966 Cert. Date: 12/20/2021 Cert. No.: 00038
Linearity File No.: 00967 Lin. Date: 12/20/2021 Lin. No.: 00042
Solution File No.: 00978 Soln. Date: 02/04/2022 Soln. No.: 00262
Sequential File No.: 00978 File Date: 02/04/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162
Control Solution %: 0.100% Expires: 06/30/2022
Solution Control Lot: 20310 Bottle No.: 1086

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	18:08S		
Control 1 EC	0.100%	18:09S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:09S		
Control 2 EC	0.100%	18:10S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:11S		
Control 3 EC	0.099%	18:11S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:12S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RUDITSKY

First Name: DAVID

MI: A

Signature: _____

Badge No.: 090

Date: 02/04/2022

