

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00890 Calib. Date: 09/18/2020 Calib. No.: 00045
Certification File No.: 00891 Cert. Date: 09/18/2020 Cert. No.: 00035
Linearity File No.: 00892 Lin. Date: 09/18/2020 Lin. No.: 00039
Solution File No.: 00926 Soln. Date: 03/08/2021 Soln. No.: 00245
Sequential File No.: 00926 File Date: 03/08/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0164
Control Solution %: 0.100% Expires: 09/02/2022
Solution Control Lot: 20390 Bottle No.: 1402

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:45S		
Control 1 EC	0.099%	07:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:47S		
Control 2 EC	0.099%	07:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:48S		
Control 3 EC	0.100%	07:49S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:50S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARRONE

First Name: KENNETH

MI: L

Signature: *Ken Marrone #069*

Badge No.: 069

Date: 03/08/2021