

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00706	Calib. Date:	01/09/2018
Certification File No.:	00707	Cert. No.:	00037
Linearity File No.:	00708	Cert. Date:	01/09/2018
Solution File No.:	00717	Lin. Date:	01/09/2018
Sequential File No.:	00717	Lin. No.:	00032
		Soln. Date:	03/14/2018
		Soln. No.:	00177
		File Date:	03/14/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	17290	Expires:	09/14/2019
		Bottle No.:	0309

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:10S		
Control 1 EC	0.098%	09:11S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:12S		
Control 2 EC	0.097%	09:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:13S		
Control 3 EC	0.096%	09:14S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: RUDITSKY

First Name: DAVID

MI: A

Signature: _____



90

Badge No.: 090

Date: 03/14/2018

