

5/27

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00965	Calib. Date:	12/20/2021
Certification File No.:	00966	Cert. Date:	12/20/2021
Linearity File No.:	00967	Lin. Date:	12/20/2021
Solution File No.:	00995	Soln. Date:	05/22/2022
Sequential File No.:	00995	File Date:	05/22/2022
		Calib. No.:	00048
		Cert. No.:	00038
		Lin. No.:	00042
		Soln. No.:	00267
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	21010	Expires:	01/13/2023
		Bottle No.:	0642

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:09D		
Control 1 EC	0.101%	12:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:11D		
Control 2 EC	0.101%	12:11D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:12D		
Control 3 EC	0.101%	12:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MORGANTE

First Name: MICHAEL

MI: A

Signature: _____



Badge No.: 128

Date: 05/22/2022

