

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00788	Calib. Date: 04/15/2019
Certification File No.:	00789	Calib. No.: 00041
Linearity File No.:	00790	Cert. Date: 04/15/2019
Solution File No.:	00796	Cert. No.: 00031
Sequential File No.:	00796	Lin. Date: 04/15/2019
		Lin. No.: 00035
		Soln. Date: 06/01/2019
		Soln. No.: 00201
		File Date: 06/01/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0162
Solution Control Lot:	18290	Expires: 09/17/2020
		Bottle No.: 0840

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:29D		
Control 1 EC	0.096%	03:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	03:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:31D		
Control 2 EC	0.097%	03:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	03:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:32D		
Control 3 EC	0.096%	03:32D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	03:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TAGLIETTA

First Name: DAVID

MI: B

Signature: _____

Sgt Taglietta #102

Badge No.: 102

Date: 06/01/2019

