

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0069  
Location: MARLBORO TWP. POLICE  
Calibration File No.: 00927 Calib. Date: 03/09/2021 Calib. No.: 00046  
Certification File No.: 00928 Cert. Date: 03/09/2021 Cert. No.: 00036  
Linearity File No.: 00929 Lin. Date: 03/09/2021 Lin. No.: 00040  
Solution File No.: 00952 Soln. Date: 06/26/2021 Soln. No.: 00255  
Sequential File No.: 00952 File Date: 06/26/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162  
Control Solution %: 0.100% Expires: 09/02/2022  
Solution Control Lot: 20390 Bottle No.: 1368

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:59D		
Control 1 EC	0.100%	17:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	17:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:00D		
Control 2 EC	0.099%	17:01D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	17:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:02D		
Control 3 EC	0.098%	17:02D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	17:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TISCHLER

First Name: MICHAEL

MI: J

Signature: \_\_\_\_\_

*SA. M. Tischler*

Badge No.: 075

Date: 06/26/2021

