

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00739	Calib. Date:	07/02/2018
Certification File No.:	00740	Cert. No.:	00038
Linearity File No.:	00741	Cert. Date:	07/02/2018
Solution File No.:	00743	Lin. No.:	00033
Sequential File No.:	00743	Soln. Date:	07/03/2018
		Soln. No.:	00183
		File Date:	07/03/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0164
Solution Control Lot:	17290	Expires:	09/14/2019
		Bottle No.:	0303

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:39D		
Control 1 EC	0.100%	12:39D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:40D		
Control 2 EC	0.099%	12:41D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:41D		
Control 3 EC	0.099%	12:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARRONE

First Name: KENNETH

MI: L

Signature: _____

CM M M #69

Badge No.: 069

Date: 07/03/2018

