

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0069  
Location: MARLBORO TWP. POLICE  
Calibration File No.: 00927 Calib. Date: 03/09/2021 Calib. No.: 00046  
Certification File No.: 00928 Cert. Date: 03/09/2021 Cert. No.: 00036  
Linearity File No.: 00929 Lin. Date: 03/09/2021 Lin. No.: 00040  
Solution File No.: 00954 Soln. Date: 07/17/2021 Soln. No.: 00256  
Sequential File No.: 00954 File Date: 07/17/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0162  
Control Solution %: 0.100% Expires: 09/02/2022  
Solution Control Lot: 20390 Bottle No.: 1378

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:24D		
Control 1 EC	0.098%	16:25D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:25D		
Control 2 EC	0.098%	16:26D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:27D		
Control 3 EC	0.098%	16:27D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	16:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: MORGANTE

First Name: MICHAEL

MI: A

Signature: \_\_\_\_\_

Badge No.: 128

Date: 07/17/2021

