

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00681	Calib. Date: 07/27/2017
Certification File No.:	00682	Calib. No.: 00036
Linearity File No.:	00683	Cert. Date: 07/27/2017
Solution File No.:	00688	Cert. No.: 00027
Sequential File No.:	00688	Lin. Date: 07/27/2017
		Lin. No.: 00031
		Soln. Date: 09/07/2017
		Soln. No.: 00167
		File Date: 09/07/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0164
Solution Control Lot:	17050	Expires: 01/30/2019
		Bottle No.: 1011

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:49D		
Control 1 EC	0.103%	13:49D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.102%	13:49D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:50D		
Control 2 EC	0.101%	13:50D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.103%	13:50D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:51D		
Control 3 EC	0.102%	13:52D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:52D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARRONE

First Name: KENNETH

MI: L

Signature: *Cpl Kenneth Marrone #69*

Badge No.: 69

Date: 09/07/2017

