

# Calibrating Unit New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00767	Calib. Date:	12/06/2018
Certification File No.:	00768	Cert. Date:	12/06/2018
Linearity File No.:	00769	Lin. Date:	12/06/2018
Solution File No.:	00771	Soln. Date:	12/10/2018
Sequential File No.:	00771	File Date:	12/10/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0162
Solution Control Lot:	17290	Expires:	09/14/2019
		Bottle No.:	0636

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:05S		
Control 1 EC	0.099%	10:06S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	10:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:06S		
Control 2 EC	0.097%	10:07S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:08S		
Control 3 EC	0.097%	10:09S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:09S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: RUDITSKY	First Name: DAVID	MI: A
Signature: 	#90	Badge No.: 090
		Date: 12/10/2018

