

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0069
Location: MARLBORO TWP. POLICE
Calibration File No.: 00681 Calib. Date: 07/27/2017 Calib. No.: 00036
Certification File No.: 00682 Cert. Date: 07/27/2017 Cert. No.: 00027
Linearity File No.: 00683 Lin. Date: 07/27/2017 Lin. No.: 00031
Solution File No.: 00701 Soln. Date: 12/18/2017 Soln. No.: 00173
Sequential File No.: 00701 File Date: 12/18/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXC S3-0164
Control Solution %: 0.100% Expires: 09/14/2019
Solution Control Lot: 17290 Bottle No.: 0188

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:25S		
Control 1 EC	0.103%	11:26S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:26S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:27S		
Control 2 EC	0.101%	11:27S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.103%	11:27S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:28S		
Control 3 EC	0.102%	11:29S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.102%	11:29S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARRONE

First Name: KENNETH

MI: L

Signature: *Kenneth Marrone #69*

Badge No.: 069

Date: 12/18/2017

