

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0069
Location:	MARLBORO TWP. POLICE	
Calibration File No.:	00890	Calib. Date: 09/18/2020
Certification File No.:	00891	Calib. No.: 00045
Linearity File No.:	00892	Cert. Date: 09/18/2020
Solution File No.:	00912	Cert. No.: 00035
Sequential File No.:	00912	Lin. Date: 09/18/2020
		Lin. No.: 00039
		Soln. Date: 12/21/2020
		Soln. No.: 00241
		File Date: 12/21/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXC S3-0164
Solution Control Lot:	20390	Expires: 09/02/2022
		Bottle No.: 1401

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:54S		
Control 1 EC	0.100%	10:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:56S		
Control 2 EC	0.100%	10:56S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:57S		
Control 3 EC	0.100%	10:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:58S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: RUDITSKY

First Name: DAVID

MI: A

Signature: _____



Badge No.: 090

Date: 12/21/2020

