

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0069
Location:	MARLBORO TWP. POLICE		
Calibration File No.:	00608	Calib. Date:	03/23/2016
Certification File No.:	00609	Cert. No.:	00032
Linearity File No.:	00610	Cert. Date:	03/23/2016
Solution File No.:	00612	Lin. Date:	03/23/2016
Sequential File No.:	00612	Lin. No.:	00027
		Soln. Date:	04/21/2016
		Soln. No.:	00139
		File Date:	04/21/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXC S3-0164
Solution Control Lot:	15D138	Expires:	04/13/2017
		Bottle No.:	0375

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:00D		
Control 1 EC	0.100%	23:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	23:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:02D		
Control 2 EC	0.101%	23:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	23:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:03D		
Control 3 EC	0.100%	23:04D	34.0°C	*** TEST PASSED ***
Control-3 IR	0.099%	23:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OHARE

First Name: BRYAN

MI: F

Signature: _____

Bryan Ohare #096

Badge No.: 96

Date: 04/21/2016

